

Bruxelles, le 12 juin 2024

Qualité des Laboratoires

RAPPORT GLOBAL EEQ HbA1c SANG FRAIS - 2024.3B

Target HbA1c = 53.1 mmol/mol (Méthode de référence NGSP/DCCT)

| INSTRUMENT | RESULT (Mean) | N | SD | CV% | Bias | Bias % |
|--------------------------------------------|---------------|----|------|------|------|--------|
| ADAMS HA-8180T IFCC mmol/mol | 59.0 | 3 | 8.7 | 14.8 | 5.9 | 11.1 |
| ADAMS HA-8180V IFCC mmol/mol | 51.8 | 22 | 1.1 | 2.2 | -1.3 | -2.4 |
| ADAMS HA-8190V IFCC mmol/mol | 53.5 | 2 | 0.7 | 1.3 | 0.4 | 0.7 |
| Afinion 2 IFCC mmol/mol | 52.0 | 1 | **** | **** | -1.1 | -2.1 |
| Alinity C IFCC mmol/mol | 50.4 | 6 | 0.6 | 1.1 | -2.7 | -5.1 |
| Capillarys 2 flex – piercing IFCC mmol/mol | 53.2 | 4 | 0.9 | 1.8 | 0.1 | 0.2 |
| Capillarys 3 Octa IFCC mmol/mol | 53.4 | 12 | 1.2 | 2.2 | 0.3 | 0.6 |
| Capillarys 3 Tera IFCC mmol/mol | 53.2 | 5 | 0.4 | 0.8 | 0.1 | 0.2 |
| Cobas 6000/8000 c501/502 IFCC mmol/mol | 53.0 | 2 | 1.4 | 2.7 | -0.1 | -0.2 |
| D-10_A1c_Program IFCC mmol/mol | 53.7 | 3 | 2.2 | 4.1 | 0.6 | 1.2 |
| D-100_A1c_Program IFCC mmol/mol | 51.8 | 8 | 1.6 | 3.0 | -1.4 | -2.6 |
| Hb NEXT IFCC mmol/mol | 53.2 | 4 | 1.4 | 2.6 | 0.1 | 0.1 |
| Minicap flex-piercing IFCC mmol/mol | 52.3 | 4 | 1.0 | 1.8 | -0.9 | -1.6 |
| TOSOH G11 IFCC mmol/mol | 53.7 | 19 | 0.6 | 1.1 | 0.6 | 1.1 |
| TOSOH G8 IFCC mmol/mol | 53.9 | 24 | 1.1 | 2.1 | 0.8 | 1.5 |

Votre rapport individuel est disponible sur le site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=28

Les coordinateurs d'enquête
 Y. Lenga

Dr. K. Vernelen