

Bruxelles, le 12 juin 2024

Qualité des Laboratoires

**RAPPORT GLOBAL EEQ HbA1c SANG FRAIS - 2024.3C**

**Target HbA1c = 44.2 mmol/mol (Méthode de référence NGSP/DCCT)**

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180T IFCC mmol/mol	45.0	3	1.0	2.2	0.8	1.9
ADAMS HA-8180V IFCC mmol/mol	43.1	22	1.1	2.6	-1.1	-2.4
ADAMS HA-8190V IFCC mmol/mol	44.5	2	0.7	1.6	0.3	0.8
Afinion 2 IFCC mmol/mol	43.0	1	****	****	-1.2	-2.6
Alinity C IFCC mmol/mol	41.4	6	0.3	0.8	-2.8	-6.3
Capillarys 2 flex – piercing IFCC mmol/mol	43.6	4	0.7	1.5	-0.6	-1.3
Capillarys 3 Octa IFCC mmol/mol	43.6	12	1.2	2.8	-0.6	-1.3
Capillarys 3 Tera IFCC mmol/mol	44.0	5	0.7	1.6	-0.2	-0.3
Cobas 6000/8000 c501/502 IFCC mmol/mol	44.0	2	1.4	3.2	-0.2	-0.3
D-10_A1c_Program IFCC mmol/mol	44.5	3	1.7	3.9	0.3	0.7
D-100_A1c_Program IFCC mmol/mol	42.3	8	1.4	3.2	-1.8	-4.1
Hb NEXT IFCC mmol/mol	44.6	4	1.6	3.5	0.4	0.9
Minicap flex-piercing IFCC mmol/mol	42.8	4	1.0	2.2	-1.4	-3.2
TOSOH G11 IFCC mmol/mol	44.5	19	0.5	1.1	0.3	0.7
TOSOH G8 IFCC mmol/mol	44.7	24	0.8	1.9	0.5	1.1

Votre rapport individuel est disponible sur le site  
[https://qseenet.com/CueSee\\_frame.asp?SelectMenu1=28](https://qseenet.com/CueSee_frame.asp?SelectMenu1=28)

Les coordinateurs d'enquête  
 Y. Lenga

Dr. K. Vernelen