

Bruxelles, le 23 février 2022

Qualité des Laboratoires

RAPPORT.GLOBAL EEQ HbA1c SANG FRAIS - 2022.1C

Target HbA1c = 55.8 mmol/mol (Méthode de référence NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Alinity C IFCC mmol/mol	56.1	3	0.2	0.3	0.3	0.5
Architect Enzymatic test C IFCC mmol/mol	56.0	1	****	****	0.2	0.3
Capillarys 2 flex – piercing IFCC mmol/mol	57.0	6	1.8	3.1	1.2	2.1
Capillarys 3 Octa IFCC mmol/mol	56.0	8	0.5	1.0	0.2	0.3
Capillarys 3 Tera IFCC mmol/mol	56.0	4	0.8	1.5	0.2	0.3
Cobas c 311/111 IFCC mmol/mol	56.3	1	****	****	0.5	0.8
Cobas 6000/8000 c501/502 IFCC mmol/mol	56.0	4	0.8	1.5	0.2	0.3
D-10_A1c_Program IFCC mmol/mol	57.7	3	4.0	7.0	1.8	3.3
D-100_A1c_Program IFCC mmol/mol	55.9	9	1.7	3.0	0.1	0.2
Menarini HA-8160 TP IFCC mmol/mol	56.0	2	1.4	2.5	0.2	0.3
Menarini HA-8160 VP IFCC mmol/mol	57.5	8	3.5	6.0	1.7	3.0
Menarini HA-8180T IFCC mmol/mol	56.8	4	1.9	3.3	0.9	1.6
Menarini HA-8180V IFCC mmol/mol	56.0	27	1.6	2.9	0.2	0.3
Minicap flex-piercing IFCC mmol/mol	55.8	5	0.8	1.5	0.0	-0.1
TOSOH G11 IFCC mmol/mol	56.1	7	1.0	1.8	0.2	0.4
TOSOH G8 IFCC mmol/mol	56.8	29	1.1	2.0	0.9	1.7
Variant-II_Turbo_A1c_Program IFCC mmol/mol	56.0	1	****	****	0.2	0.3

Votre rapport individuel est disponible sur le site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=28

Les coordinateurs d'enquête
 Y. Lenga

Dr. K. Vernelen