

Bruxelles, le 23 février 2022

Qualité des Laboratoires

**RAPPORT.GLOBAL EEQ HbA1c SANG FRAIS - 2022.1D**

**Target HbA1c = 68.5 mmol/mol (Méthode de référence NGSP/DCCT)**

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Alinity C IFCC mmol/mol	69.5	3	0.5	0.7	0.9	1.4
Architect Enzymatic test C IFCC mmol/mol	69.0	1	****	****	0.5	0.7
Capillarys 2 flex – piercing IFCC mmol/mol	68.3	6	1.6	2.4	-0.2	-0.3
Capillarys 3 Octa IFCC mmol/mol	68.0	8	1.5	2.2	-0.5	-0.8
Capillarys 3 Tera IFCC mmol/mol	68.5	4	1.3	1.9	0.0	0.0
Cobas c 311/111 IFCC mmol/mol	70.0	1	****	****	1.5	2.2
Cobas 6000/8000 c501/502 IFCC mmol/mol	69.8	4	0.5	0.7	1.2	1.8
D-10_A1c_Program IFCC mmol/mol	70.0	3	5.3	7.6	1.5	2.2
D-100_A1c_Program IFCC mmol/mol	68.6	9	1.2	1.8	0.1	0.1
Menarini HA-8160 TP IFCC mmol/mol	68.5	2	2.1	3.1	0.0	0.0
Menarini HA-8160 VP IFCC mmol/mol	70.0	8	4.1	5.9	1.5	2.2
Menarini HA-8180T IFCC mmol/mol	70.0	4	2.1	3.1	1.5	2.2
Menarini HA-8180V IFCC mmol/mol	68.4	27	2.0	2.9	-0.2	-0.2
Minicap flex-piercing IFCC mmol/mol	68.0	5	1.0	1.5	-0.5	-0.8
TOSOH G11 IFCC mmol/mol	70.2	7	1.1	1.6	1.6	2.4
TOSOH G8 IFCC mmol/mol	70.5	29	1.2	1.7	2.0	2.9
Variant-II_Turbo_A1c_Program IFCC mmol/mol	68.5	1	****	****	0.0	0.0

Votre rapport individuel est disponible sur le site  
[https://qseenet.com/CueSee\\_frame.asp?SelectMenu1=28](https://qseenet.com/CueSee_frame.asp?SelectMenu1=28)

Les coordinateurs d'enquête  
 Y. Lenga

Dr. K. Vernelen