

Bruxelles, le 09 novembre 2022

Qualité des Laboratoires

RAPPORT GLOBAL EEQ HbA1c SANG FRAIS - 2022.6C

Target HbA1c = 36.7 mmol/mol (Méthode de référence NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Alinity C IFCC mmol/mol	35.5	6	1.5	4.3	-1.2	-3.2
Capillarys 2 flex – piercing IFCC mmol/mol	36.8	6	0.8	2.0	0.1	0.3
Capillarys 3 Octa IFCC mmol/mol	36.0	10	0.8	2.3	-0.7	-2.0
Capillarys 3 Tera IFCC mmol/mol	36.4	8	0.7	2.0	-0.3	-0.9
Cobas c 303 IFCC mmol/mol (Cobas pure)	39.0	1	****	****	2.3	6.2
Cobas 6000/8000 c501/502 IFCC mmol/mol	37.0	2	1.4	3.8	0.3	0.8
D-10_A1c_Program IFCC mmol/mol	38.5	2	0.7	1.8	1.8	4.8
D-100_A1c_Program IFCC mmol/mol	36.3	9	0.5	1.3	-0.4	-1.1
Menarini HA-8160 TP IFCC mmol/mol	36.0	1	****	****	-0.7	-2.0
Menarini HA-8160 VP IFCC mmol/mol	36.0	1	****	****	-0.7	-2.0
Menarini HA-8180T IFCC mmol/mol	37.3	4	0.5	1.3	0.5	1.4
Menarini HA-8180V IFCC mmol/mol	37.0	23	1.3	3.5	0.3	0.7
Menarini HA-8190V IFCC mmol/mol	36.5	2	0.7	1.9	-0.2	-0.6
Menarini Hb NEXT IFCC mmol/mol	38.3	3	1.7	4.6	1.5	4.2
Minicap flex-piercing IFCC mmol/mol	36.5	4	0.6	1.6	-0.2	-0.6
TOSOH G11 IFCC mmol/mol	37.7	11	0.4	1.1	1.0	2.7
TOSOH G8 IFCC mmol/mol	38.3	30	0.8	2.0	1.6	4.4

Votre rapport individuel est disponible sur le site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=28

Les coordinateurs d'enquête
 Y. Lenga

Dr. K. Vernelen