

Bruxelles, le 09 novembre 2022

Qualité des Laboratoires

RAPPORT GLOBAL EEQ HbA1c SANG FRAIS - 2022.6D

Target HbA1c = 74.5 mmol/mol (Méthode de référence NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Alinity C IFCC mmol/mol	73.9	6	2.2	3.1	-0.7	-0.9
Capillarys 2 flex – piercing IFCC mmol/mol	75.3	6	1.6	2.2	0.8	1.1
Capillarys 3 Octa IFCC mmol/mol	74.5	10	1.3	1.7	0.0	0.0
Capillarys 3 Tera IFCC mmol/mol	75.0	8	0.9	1.2	0.5	0.6
Cobas c 303 IFCC mmol/mol (Cobas pure)	78.0	1	****	****	3.5	4.6
Cobas c 311/111 IFCC mmol/mol	69.8	1	****	****	-4.7	-6.4
Cobas 6000/8000 c501/502 IFCC mmol/mol	72.0	2	4.2	5.9	-2.5	-3.4
D-10_A1c_Program IFCC mmol/mol	77.0	2	1.4	1.8	2.5	3.3
D-100_A1c_Program IFCC mmol/mol	73.1	9	1.5	2.1	-1.4	-1.9
Menarini HA-8160 TP IFCC mmol/mol	74.0	1	****	****	-0.5	-0.7
Menarini HA-8160 VP IFCC mmol/mol	74.0	1	****	****	-0.5	-0.7
Menarini HA-8180T IFCC mmol/mol	74.3	4	1.3	1.7	-0.3	-0.4
Menarini HA-8180V IFCC mmol/mol	74.4	23	2.1	2.9	-0.1	-0.1
Menarini HA-8190V IFCC mmol/mol	75.5	2	0.7	0.9	1.0	1.3
Menarini Hb NEXT IFCC mmol/mol	77.5	3	6.5	8.4	3.0	4.0
Minicap flex-piercing IFCC mmol/mol	75.0	4	1.8	2.4	0.5	0.6
TOSOH G11 IFCC mmol/mol	74.9	11	0.5	0.7	0.3	0.4
TOSOH G8 IFCC mmol/mol	75.5	29	1.0	1.4	1.0	1.3

Votre rapport individuel est disponible sur le site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=28

Les coordinateurs d'enquête
 Y. Lenga

Dr. K. Vernelen