

Bruxelles, le 31 janvier 2023

Qualité des Laboratoires

RAPPORT GLOBAL EEQ HbA1c SANG FRAIS - 2023.1B

Target HbA1c = 51.3 mmol/mol (Méthode de référence NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8160 TP IFCC mmol/mol	50.0	1	*****	*****	-1.3	-2.5
ADAMS HA-8160 VP IFCC mmol/mol	50.0	2	4.2	8.4	-1.3	-2.5
ADAMS HA-8180T IFCC mmol/mol	51.0	3	1.0	2.0	-0.3	-0.5
ADAMS HA-8180V IFCC mmol/mol	50.1	21	1.8	3.7	-1.2	-2.3
ADAMS HA-8190V IFCC mmol/mol	51.0	2	0.0	0.0	-0.3	-0.5
Afinion AS100 IFCC mmol/mol	49.0	1	*****	*****	-2.3	-4.4
Alinity C IFCC mmol/mol	51.1	6	0.9	1.8	-0.2	-0.4
Capillarys 2 flex – piercing IFCC mmol/mol	50.8	6	0.7	1.4	-0.5	-0.9
Capillarys 3 Octa IFCC mmol/mol	50.9	11	0.5	1.1	-0.3	-0.7
Capillarys 3 Tera IFCC mmol/mol	51.0	6	1.1	2.1	-0.3	-0.5
Cobas c 303 IFCC mmol/mol (cobas pure)	54.0	1	*****	*****	2.7	5.4
Cobas 6000/8000 c501/502 IFCC mmol/mol	51.5	2	0.7	1.4	0.2	0.5
D-10_A1c_Program IFCC mmol/mol	52.0	1	*****	*****	0.7	1.4
D-100_A1c_Program IFCC mmol/mol	48.5	10	0.7	1.4	-2.8	-5.5
Hb NEXT IFCC mmol/mol	54.3	3	5.9	10.8	3.1	6.0
Minicap flex-piercing IFCC mmol/mol	50.0	2	2.8	5.7	-1.3	-2.5
TOSOH G11 IFCC mmol/mol	50.7	12	0.6	1.2	-0.6	-1.2
TOSOH G8 IFCC mmol/mol	51.5	30	1.1	2.1	0.2	0.5

Votre rapport individuel est disponible sur le site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=28

Les coordinateurs d'enquête
 Y. Lenga

Dr. K. Vernelen