

Bruxelles, le 31 janvier 2023

Qualité des Laboratoires

RAPPORT GLOBAL EEQ HbA1c SANG FRAIS - 2023.1D

Target HbA1c = 51.3 mmol/mol (Méthode de référence NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8160 TP IFCC mmol/mol	49.0	1	*****	*****	-2.3	-4.4
ADAMS HA-8160 VP IFCC mmol/mol	59.5	2	9.3	16.0	8.2	16.0
ADAMS HA-8180T IFCC mmol/mol	51.0	3	1.0	2.0	-0.3	-0.5
ADAMS HA-8180V IFCC mmol/mol	50.8	21	2.1	4.2	-0.5	-0.9
ADAMS HA-8190V IFCC mmol/mol	51.5	2	0.7	1.4	0.2	0.5
Afinion AS100 IFCC mmol/mol	51.0	1	*****	*****	-0.3	-0.5
Alinity C IFCC mmol/mol	52.1	6	0.8	1.5	0.8	1.6
Capillarys 2 flex – piercing IFCC mmol/mol	51.3	6	0.8	1.5	0.0	0.1
Capillarys 3 Octa IFCC mmol/mol	50.4	11	0.9	1.8	-0.9	-1.7
Capillarys 3 Tera IFCC mmol/mol	54.2	6	8.8	16.0	2.9	5.7
Cobas c 303 IFCC mmol/mol (cobas pure)	53.0	1	*****	*****	1.7	3.4
Cobas 6000/8000 c501/502 IFCC mmol/mol	53.5	2	0.7	1.3	2.2	4.4
D-10_A1c_Program IFCC mmol/mol	52.5	2	0.7	1.3	1.2	2.4
D-100_A1c_Program IFCC mmol/mol	48.8	10	0.6	1.2	-2.4	-4.8
Hb NEXT IFCC mmol/mol	55.0	3	4.8	8.7	3.8	7.4
Minicap flex-piercing IFCC mmol/mol	47.3	3	6.4	13.0	-3.9	-7.7
TOSOH G11 IFCC mmol/mol	50.8	12	0.6	1.2	-0.4	-0.9
TOSOH G8 IFCC mmol/mol	51.6	30	0.8	1.6	0.3	0.6

Votre rapport individuel est disponible sur le site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=28

Les coordinateurs d'enquête
 Y. Lenga

Dr. K. Vernelen