

Bruxelles, le 30 mai 2023

Qualité des Laboratoires

**RAPPORT GLOBAL EEQ HbA1c SANG FRAIS - 2023.3C**

**Target HbA1c = 33.2 mmol/mol (Méthode de référence NGSP/DCCT)**

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180T IFCC mmol/mol	33.0	3	1.7	2.3	-1.8	-2.4
ADAMS HA-8180V IFCC mmol/mol	33.0	20	1.6	2.2	-1.6	-2.1
ADAMS HA-8190V IFCC mmol/mol	32.5	2	0.0	0.0	-0.8	-1.1
Alinity C IFCC mmol/mol	31.4	6	0.7	0.9	-1.8	-2.4
Capillarys 2 flex – piercing IFCC mmol/mol	32.6	6	1.3	1.8	-1.8	-2.4
Capillarys 3 Octa IFCC mmol/mol	32.8	12	1.0	1.3	-1.9	-2.5
Capillarys 3 Tera IFCC mmol/mol	33.3	7	1.0	1.4	-1.8	-2.4
Cobas 6000/8000 c501/502 IFCC mmol/mol	31.0	2	0.0	0.0	-0.8	-1.1
D-10_A1c_Program IFCC mmol/mol	34.0	2	2.1	2.7	1.7	2.2
D-100_A1c_Program IFCC mmol/mol	30.2	8	2.3	3.2	-2.5	-3.2
Hb NEXT IFCC mmol/mol	31.5	3	4.1	5.4	0.9	1.2
Minicap flex-piercing IFCC mmol/mol	35.3	3	6.7	8.4	3.5	4.6
TOSOH G11 IFCC mmol/mol	32.2	18	1.0	1.3	-1.1	-1.4
TOSOH G8 IFCC mmol/mol	32.7	24	0.8	1.0	-0.3	-0.4

Votre rapport individuel est disponible sur le site  
[https://gseenet.com/CueSee\\_frame.asp?SelectMenu1=28](https://gseenet.com/CueSee_frame.asp?SelectMenu1=28)

Les coordinateurs d'enquête  
 Y. Lenga

Dr. K. Vernelen