

Bruxelles, le 30 mai 2023

Qualité des Laboratoires

RAPPORT GLOBAL EEQ HbA1c SANG FRAIS - 2023.3D

Target HbA1c = 48.2 mmol/mol (Méthode de référence NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180T IFCC mmol/mol	46.3	3	0.6	1.2	-1.9	-3.9
ADAMS HA-8180V IFCC mmol/mol	46.8	20	1.2	2.6	-1.4	-2.8
ADAMS HA-8190V IFCC mmol/mol	46.5	2	0.7	1.5	-1.7	-3.5
Alinity C IFCC mmol/mol	46.2	6	0.5	1.2	-2.0	-4.2
Capillarys 2 flex – piercing IFCC mmol/mol	47.0	6	1.2	2.5	-1.2	-2.6
Capillarys 3 Octa IFCC mmol/mol	47.6	12	1.0	2.1	-0.6	-1.3
Capillarys 3 Tera IFCC mmol/mol	47.7	7	1.1	2.3	-0.5	-1.0
Cobas 6000/8000 c501/502 IFCC mmol/mol	49.0	2	0.0	0.0	0.8	1.7
D-10_A1c_Program IFCC mmol/mol	49.0	2	1.4	2.9	0.8	1.7
D-100_A1c_Program IFCC mmol/mol	44.1	8	0.9	2.0	-4.1	-8.6
Hb NEXT IFCC mmol/mol	49.2	3	2.0	4.0	1.0	2.1
Minicap flex-piercing IFCC mmol/mol	49.7	3	2.9	5.8	1.5	3.0
TOSOH G11 IFCC mmol/mol	48.0	18	0.6	1.2	-0.2	-0.5
TOSOH G8 IFCC mmol/mol	48.6	24	0.5	1.1	0.4	0.8

Votre rapport individuel est disponible sur le site
https://gseenet.com/CueSee_frame.asp?SelectMenu1=28

Les coordinateurs d'enquête
 Y. Lenga

Dr. K. Vernelen