

Bruxelles, le 07 novembre 2023

Qualité des Laboratoires

RAPPORT GLOBAL EEQ HbA1c SANG FRAIS - 2023.6B

Target HbA1c = 59.7 mmol/mol (Méthode de référence NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8160VP IFCC mmol/mol	60.3	2	0.4	0.7	0.6	1.1
ADAMS HA-8180T IFCC mmol/mol	60.3	3	0.6	1.0	0.7	1.1
ADAMS HA-8180V IFCC mmol/mol	58.1	21	1.6	2.7	-1.5	-2.6
ADAMS HA-8190V IFCC mmol/mol	61.0	2	1.4	2.3	1.3	2.2
Afinion AS100 IFCC mmol/mol	59.0	1	****	****	-0.7	-1.1
Alinity C IFCC mmol/mol	58.1	6	0.9	1.5	-1.6	-2.7
Capillarys 2 flex – piercing IFCC mmol/mol	58.5	7	2.5	4.2	-1.1	-1.9
Capillarys 3 Octa IFCC mmol/mol	58.8	13	1.4	2.3	-0.9	-1.5
Capillarys 3 Tera IFCC mmol/mol	58.9	7	1.7	2.8	-0.8	-1.4
Cobas 6000/8000 c501/502 IFCC mmol/mol	60.5	2	0.7	1.2	0.8	1.4
D-10_A1c_Program IFCC mmol/mol	60.0	2	0.0	0.0	0.3	0.5
D-100_A1c_Program IFCC mmol/mol	57.8	8	1.0	1.7	-1.9	-3.1
Hb NEXT IFCC mmol/mol	61.3	4	1.0	1.7	1.6	2.6
Minicap flex-piercing IFCC mmol/mol	57.2	4	3.0	5.2	-2.4	-4.1
TOSOH G11 IFCC mmol/mol	59.8	18	0.8	1.3	0.1	0.2
TOSOH G8 IFCC mmol/mol	60.6	26	1.1	1.8	0.9	1.5

Votre rapport individuel est disponible sur le site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=28

Les coordinateurs d'enquête
 Y. Lenga

Dr. K. Vernelen