

Brussel, 23 februari 2024

**Kwaliteit van Laboratoria**

**GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2024.1A**

**Target HbA1c = 51.8 mmol/mol (Referentiemethode NGSP/DCCT)**

<b>INSTRUMENT</b>	<b>RESULT (Mean)</b>	<b>N</b>	<b>SD</b>	<b>CV%</b>	<b>Bias</b>	<b>Bias %</b>
ADAMS HA-8180T IFCC mmol/mol	53.0	3	0.0	0.0	1.2	2.3
ADAMS HA-8180V IFCC mmol/mol	51.2	20	1.8	3.5	-0.6	-1.2
ADAMS HA-8190V IFCC mmol/mol	53.0	2	0.0	0.0	1.2	2.3
Afinion AS100 IFCC mmol/mol	48.0	1	****	****	-3.8	-7.3
Alinity C IFCC mmol/mol	50.5	6	0.6	1.3	-1.3	-2.6
Capillarys 2 flex – piercing IFCC mmol/mol	50.8	5	1.7	3.3	-1.0	-1.9
Capillarys 3 Octa IFCC mmol/mol	50.6	13	0.9	1.8	-1.2	-2.3
Capillarys 3 Tera IFCC mmol/mol	51.6	7	2.4	4.7	-0.2	-0.4
Cobas 6000/8000 c501/502 IFCC mmol/mol	53.5	2	2.1	4.0	1.7	3.3
D-10_A1c_Program IFCC mmol/mol	53.7	3	1.2	2.3	1.9	3.7
D-100_A1c_Program IFCC mmol/mol	50.4	8	1.3	2.5	-1.4	-2.8
Hb NEXT IFCC mmol/mol	51.7	6	1.1	2.1	-0.1	-0.1
Minicap flex-piercing IFCC mmol/mol	50.3	3	1.5	3.0	-1.5	-2.8
TOSOH G11 IFCC mmol/mol	52.1	19	1.0	1.9	0.3	0.6
TOSOH G8 IFCC mmol/mol	52.3	24	2.6	5.0	0.5	0.9

Uw individueel rapport kunt u opvragen via de site  
[https://qseenet.com/CueSee\\_frame.asp?SelectMenu1=2](https://qseenet.com/CueSee_frame.asp?SelectMenu1=2)

De enquêtecoördinatoren  
 Y. Lenga

Dr. K. Vernelen