

Brussel, 23 februari 2024

Kwaliteit van Laboratoria

GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2024.1B

Target HbA1c = 50.1 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180T IFCC mmol/mol	50.7	3	0.6	1.1	0.6	1.2
ADAMS HA-8180V IFCC mmol/mol	49.1	20	1.9	3.9	-0.9	-1.9
ADAMS HA-8190V IFCC mmol/mol	51.0	2	0.0	0.0	0.9	1.9
Afinion AS100 IFCC mmol/mol	52.0	1	****	****	1.9	3.9
Alinity C IFCC mmol/mol	48.0	6	0.6	1.3	-2.1	-4.1
Capillarys 2 flex – piercing IFCC mmol/mol	49.5	5	1.5	3.0	-0.5	-1.1
Capillarys 3 Octa IFCC mmol/mol	49.5	13	0.9	1.8	-0.5	-1.1
Capillarys 3 Tera IFCC mmol/mol	49.9	7	1.3	2.7	-0.2	-0.4
Cobas 6000/8000 c501/502 IFCC mmol/mol	50.5	2	2.1	4.2	0.4	0.9
D-10_A1c_Program IFCC mmol/mol	50.3	3	0.5	0.9	0.2	0.4
D-100_A1c_Program IFCC mmol/mol	47.8	8	0.7	1.4	-2.2	-4.5
Hb NEXT IFCC mmol/mol	49.5	6	1.0	2.1	-0.6	-1.2
Minicap flex-piercing IFCC mmol/mol	49.0	3	2.0	4.1	-1.1	-2.1
TOSOH G11 IFCC mmol/mol	49.4	19	0.8	1.5	-0.6	-1.3
TOSOH G8 IFCC mmol/mol	50.3	24	1.7	3.3	0.2	0.5

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu=2

De enquêtecoördinatoren
 Y. Lenga

Dr. K. Vernelen