

Brussel, 23 februari 2024

Kwaliteit van Laboratoria

GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2024.1D

Target HbA1c = 50.1 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180T IFCC mmol/mol	50.7	3	0.6	1.1	0.6	1.2
ADAMS HA-8180V IFCC mmol/mol	49.7	20	1.6	3.3	-0.4	-0.7
ADAMS HA-8190V IFCC mmol/mol	51.0	2	0.0	0.0	0.9	1.9
Afinion AS100 IFCC mmol/mol	49.0	1	****	****	-1.1	-2.1
Alinity C IFCC mmol/mol	48.3	6	0.7	1.5	-1.8	-3.6
Capillarys 2 flex – piercing IFCC mmol/mol	49.3	5	1.1	2.2	-0.8	-1.5
Capillarys 3 Octa IFCC mmol/mol	49.0	13	1.0	2.1	-1.0	-2.0
Capillarys 3 Tera IFCC mmol/mol	49.6	5	0.9	1.8	-0.5	-0.9
Cobas 6000/8000 c501/502 IFCC mmol/mol	52.0	2	2.8	5.4	1.9	3.9
D-10_A1c_Program IFCC mmol/mol	49.9	3	0.9	1.8	-0.1	-0.2
D-100_A1c_Program IFCC mmol/mol	47.1	8	1.2	2.5	-2.9	-5.8
Hb NEXT IFCC mmol/mol	48.8	6	4.7	9.7	-1.3	-2.6
Minicap flex-piercing IFCC mmol/mol	49.0	3	1.7	3.5	-1.1	-2.1
TOSOH G11 IFCC mmol/mol	49.5	19	0.6	1.2	-0.5	-1.0
TOSOH G8 IFCC mmol/mol	50.4	24	1.2	2.3	0.4	0.7

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=2

De enquêtecoördinatoren
 Y. Lenga

Dr. K. Vernelen