

Brussel, 27 januari 2025

**Kwaliteit van Laboratoria**

**GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2024.6B**

**Target HbA1c = 45.2 mmol/mol (Referentiemethode NGSP/DCCT)**

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180T IFCC mmol/mol	45.5	2	0.70	1.6	0.3	0.6
ADAMS HA-8180V IFCC mmol/mol	44.7	22	1.48	3.3	-0.5	-1.1
ADAMS HA-8190V IFCC mmol/mol	46.0	1	****	****	0.8	1.7
Afinion 2 IFCC mmol/mol	45.0	1	****	****	-0.2	-0.5
Afinion AS100 IFCC mmol/mol	45.0	1	****	****	-0.2	-0.5
Alinity C IFCC mmol/mol	43.1	1	****	****	-2.1	-4.7
Capillarys 2 flex-piercing IFCC mmol/mol	44.4	5	1.67	3.8	-0.8	-1.9
Capillarys 3 Octa IFCC mmol/mol	43.7	9	1.00	2.3	-1.6	-3.5
Capillarys 3 Tera IFCC mmol/mol	44.3	6	0.82	1.8	-0.9	-2.0
cobas b 101 IFCC mmol/mol	44.0	1	****	****	-1.2	-2.8
cobas c 501/502 cobas 6000 8000 IFCC	43.0	2	4.24	9.9	-2.2	-5.0
D10 A1c IFCC mmol/mol	46.7	2	0.42	0.9	1.5	3.2
D100 A1c IFCC mmol/mol	44.0	7	1.04	2.4	-1.2	-2.7
Hb NEXT IFCC mmol/mol	46.7	5	1.45	3.1	1.5	3.2
Minicap flex-piercing IFCC mmol/mol	43.5	4	1.91	4.4	-1.7	-3.9
TOSOH G11 IFCC mmol/mol	46.2	27	0.60	1.3	1.0	2.2
TOSOH G8 IFCC mmol/mol	46.4	16	0.84	1.8	1.2	2.6

Uw individueel rapport kunt u opvragen via de site  
[https://qseenet.com/CueSee\\_frame.asp?SelectMenu1=2](https://qseenet.com/CueSee_frame.asp?SelectMenu1=2)

De enquêtecoördinatoren  
 Y. Lenga

Dr. K. Vernelen/ A. Vantorre