

Brussel, 31 januari 2023

Kwaliteit van Laboratoria

GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2023.1A

Target HbA1c = 36.6 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8160 TP IFCC mmol/mol	36.0	1	*****	*****	-0.6	-1.7
ADAMS HA-8160 VP IFCC mmol/mol	35.8	2	2.5	7.1	-0.8	-2.2
ADAMS HA-8180T IFCC mmol/mol	36.5	4	1.3	3.5	-0.1	-0.3
ADAMS HA-8180V IFCC mmol/mol	36.3	21	1.3	3.5	-0.3	-0.9
ADAMS HA-8190V IFCC mmol/mol	36.5	2	0.7	1.9	-0.1	-0.3
Afinion AS100 IFCC mmol/mol	35.0	1	*****	*****	-1.6	-4.4
Alinity C IFCC mmol/mol	36.1	6	1.2	3.2	-0.5	-1.4
Capillarys 2 flex – piercing IFCC mmol/mol	37.1	6	1.0	2.8	0.5	1.3
Capillarys 3 Octa IFCC mmol/mol	36.4	11	0.8	2.2	-0.2	-0.7
Capillarys 3 Tera IFCC mmol/mol	37.0	6	1.1	3.0	0.4	1.1
Cobas c 303 IFCC mmol/mol (cobas pure)	38.0	1	*****	*****	1.4	3.8
Cobas 6000/8000 c501/502 IFCC mmol/mol	37.0	2	1.4	3.8	0.4	1.1
D-10_A1c_Program IFCC mmol/mol	38.5	2	0.7	1.8	1.9	5.2
D-100_A1c_Program IFCC mmol/mol	35.1	10	0.9	2.5	-1.6	-4.2
Hb NEXT IFCC mmol/mol	38.3	3	4.4	11.5	1.7	4.6
Minicap flex-piercing IFCC mmol/mol	45.0	3	14.7	32.7*	8.4	22.9
TOSOH G11 IFCC mmol/mol	37.1	12	0.4	1.0	0.5	1.4
TOSOH G8 IFCC mmol/mol	37.5	30	0.7	2.0	0.8	2.3

*Individuele resultaten van methode Minicap flex piercing IFCC mmol/mol (CV = 32.7%) : 37.0, 36.0 en 62.0 mmol/mol.

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu=2

De enquêtecoördinatoren
 Y. Lenga

Dr. K. Vernelen