

Brussel, 31 januari 2023

Kwaliteit van Laboratoria

GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2023.1C

Target HbA1c = 73.7 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8160 TP IFCC mmol/mol	71.0	1	*****	*****	-2.7	-3.6
ADAMS HA-8160 VP IFCC mmol/mol	61.9	2	18.2	29.5*	-11.8	-16.0
ADAMS HA-8180T IFCC mmol/mol	71.3	3	1.1	1.6	-2.3	-3.2
ADAMS HA-8180V IFCC mmol/mol	70.3	21	3.2	4.5	-3.3	-4.5
ADAMS HA-8190V IFCC mmol/mol	72.0	2	0.0	0.0	-1.7	-2.3
Afinion AS100 IFCC mmol/mol	75.0	1	*****	*****	1.3	1.8
Alinity C IFCC mmol/mol	73.2	6	0.9	1.2	-0.5	-0.7
Capillarys 2 flex – piercing IFCC mmol/mol	70.8	6	1.2	1.7	-2.9	-4.0
Capillarys 3 Octa IFCC mmol/mol	70.6	11	0.9	1.3	-3.0	-4.1
Capillarys 3 Tera IFCC mmol/mol	66.7	6	7.3	10.9	-7.0	-9.5
Cobas c 303 IFCC mmol/mol (cobas pure)	74.0	1	*****	*****	0.3	0.5
Cobas 6000/8000 c501/502 IFCC mmol/mol	73.5	2	0.7	1.0	-0.2	-0.2
D-10_A1c_Program IFCC mmol/mol	73.5	2	0.7	1.0	-0.2	-0.2
D-100_A1c_Program IFCC mmol/mol	69.7	10	1.4	2.1	-4.0	-5.4
Hb NEXT IFCC mmol/mol	76.9	3	5.3	6.8	3.2	4.4
Minicap flex-piercing IFCC mmol/mol	67.7	3	6.7	9.8	-6.0	-8.1
TOSOH G11 IFCC mmol/mol	71.8	12	0.7	0.9	-1.9	-2.6
TOSOH G8 IFCC mmol/mol	72.7	30	0.9	1.3	-1.0	-1.4

*Individuele resultaten van methode ADAMS HA-8160 VP IFCC mmol/mol (CV = 29.5%) : 74.8 en 49.0 mmol/mol.

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu=2

De enquêtecöördinatoren
 Y. Lenga

Dr. K. Vernelen