

Brussel, 30 mei 2023

**Kwaliteit van Laboratoria**

**GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2023.3A**

**Target HbA1c = 48.3 mmol/mol (Referentiemethode NGSP/DCCT)**

<b>INSTRUMENT</b>	<b>RESULT (Mean)</b>	<b>N</b>	<b>SD</b>	<b>CV%</b>	<b>Bias</b>	<b>Bias %</b>
ADAMS HA-8180T IFCC mmol/mol	46.3	3	0.6	1.2	-2.0	-4.1
ADAMS HA-8180V IFCC mmol/mol	47.6	21	1.2	2.4	-0.7	-1.6
ADAMS HA-8190V IFCC mmol/mol	48.0	2	0.0	0.0	-0.3	-0.6
Alinity C IFCC mmol/mol	47.2	6	0.7	1.5	-1.1	-2.3
Capillarys 2 flex – piercing IFCC mmol/mol	47.5	6	1.0	2.2	-0.8	-1.7
Capillarys 3 Octa IFCC mmol/mol	48.2	12	2.3	4.8	-0.1	-0.3
Capillarys 3 Tera IFCC mmol/mol	48.0	7	1.0	2.1	-0.3	-0.6
Cobas 6000/8000 c501/502 IFCC mmol/mol	49.5	2	0.7	1.4	1.2	2.5
D-10_A1c_Program IFCC mmol/mol	49.0	2	1.4	2.9	0.7	1.4
D-100_A1c_Program IFCC mmol/mol	44.3	9	0.9	2.1	-4.0	-8.3
Hb NEXT IFCC mmol/mol	50.1	4	1.4	2.8	1.8	3.8
Minicap flex-piercing IFCC mmol/mol	50.3	3	4.9	9.8	2.0	4.2
TOSOH G11 IFCC mmol/mol	47.9	18	0.8	1.7	-0.4	-0.9
TOSOH G8 IFCC mmol/mol	48.5	25	1.0	2.1	0.2	0.4

Uw individueel rapport kunt u opvragen via de site  
[https://qseenet.com/CueSee\\_frame.asp?SelectMenu1=2](https://qseenet.com/CueSee_frame.asp?SelectMenu1=2)

De enquêtecoördinatoren

Y. Lenga

Dr. K. Vernelen