

Brussel, 30 mei 2023

**Kwaliteit van Laboratoria**

**GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2023.3B**

**Target HbA1c = 75.8 mmol/mol (Referentiemethode NGSP/DCCT)**

<b>INSTRUMENT</b>	<b>RESULT (Mean)</b>	<b>N</b>	<b>SD</b>	<b>CV%</b>	<b>Bias</b>	<b>Bias %</b>
ADAMS HA-8180T IFCC mmol/mol	74.0	3	1.7	2.3	-1.8	-2.4
ADAMS HA-8180V IFCC mmol/mol	74.2	20	1.6	2.2	-1.6	-2.1
ADAMS HA-8190V IFCC mmol/mol	75.0	2	0.0	0.0	-0.8	-1.1
Alinity C IFCC mmol/mol	74.0	6	0.7	0.9	-1.8	-2.4
Capillarys 2 flex – piercing IFCC mmol/mol	74.0	6	1.3	1.8	-1.8	-2.4
Capillarys 3 Octa IFCC mmol/mol	73.9	12	1.0	1.3	-1.9	-2.5
Capillarys 3 Tera IFCC mmol/mol	74.0	7	1.0	1.4	-1.8	-2.4
Cobas 6000/8000 c501/502 IFCC mmol/mol	75.0	2	0.0	0.0	-0.8	-1.1
D-10_A1c_Program IFCC mmol/mol	77.5	2	2.1	2.7	1.7	2.2
D-100_A1c_Program IFCC mmol/mol	73.4	8	2.3	3.2	-2.5	-3.2
Hb NEXT IFCC mmol/mol	76.8	3	4.1	5.4	0.9	1.2
Minicap flex-piercing IFCC mmol/mol	79.3	3	6.7	8.4	3.5	4.6
TOSOH G11 IFCC mmol/mol	74.8	18	1.0	1.3	-1.1	-1.4
TOSOH G8 IFCC mmol/mol	75.5	24	0.8	1.0	-0.3	-0.4

Uw individueel rapport kunt u opvragen via de site  
[https://qseenet.com/CueSee\\_frame.asp?SelectMenu=2](https://qseenet.com/CueSee_frame.asp?SelectMenu=2)

De enquêtecoördinatoren

Y. Lenga

Dr. K. Vernelen