

Brussel, 07 november 2023

Kwaliteit van Laboratoria

GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2023.6A

Target HbA1c = 44.0 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8160VP IFCC mmol/mol	45.2	2	1.7	3.8	1.2	2.6
ADAMS HA-8180T IFCC mmol/mol	44.7	3	0.6	1.3	0.6	1.4
ADAMS HA-8180V IFCC mmol/mol	43.3	21	1.1	2.6	-0.8	-1.8
ADAMS HA-8190V IFCC mmol/mol	45.0	2	0.0	0.0	1.0	2.2
Afinion AS100 IFCC mmol/mol	42.0	1	****	****	-2.0	-4.6
Alinity C IFCC mmol/mol	42.7	6	0.8	1.9	-1.3	-2.9
Capillarys 2 flex – piercing IFCC mmol/mol	43.1	7	2.0	4.7	-1.0	-2.2
Capillarys 3 Octa IFCC mmol/mol	43.2	13	0.6	1.3	-0.9	-2.0
Capillarys 3 Tera IFCC mmol/mol	43.6	7	1.0	2.2	-0.5	-1.1
Cobas 6000/8000 c501/502 IFCC mmol/mol	44.0	2	0.0	0.0	0.0	-0.1
D-10_A1c_Program IFCC mmol/mol	44.5	2	0.7	1.6	0.5	1.0
D-100_A1c_Program IFCC mmol/mol	42.9	8	0.8	1.9	-1.2	-2.6
Hb NEXT IFCC mmol/mol	44.6	4	1.4	3.2	0.5	1.2
Minicap flex-piercing IFCC mmol/mol	43.0	4	1.4	3.3	-1.0	-2.4
TOSOH G11 IFCC mmol/mol	44.8	18	0.7	1.6	0.8	1.8
TOSOH G8 IFCC mmol/mol	45.1	26	0.8	1.7	1.1	2.4

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu=2

De enquêtecoördinatoren

Y. Lenga

Dr. K. Vernelen