

Brussel, 12 juni 2024

**Kwaliteit van Laboratoria**

**GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2024.3A**

**Target HbA1c = 66.4 mmol/mol (Referentiemethode NGSP/DCCT)**

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180T IFCC mmol/mol	63.3	3	7.2	11.4	-3.1	-4.7
ADAMS HA-8180V IFCC mmol/mol	65.5	22	1.7	2.5	-1.0	-1.5
ADAMS HA-8190V IFCC mmol/mol	67.0	2	0.0	0.0	0.6	0.8
Afinion AS100 IFCC mmol/mol	65.0	1	****	****	-1.4	-2.2
Alinity C IFCC mmol/mol	64.5	6	0.5	0.8	-1.9	-2.9
Capillarys 2 flex – piercing IFCC mmol/mol	66.2	4	1.1	1.7	-0.3	-0.4
Capillarys 3 Octa IFCC mmol/mol	66.6	12	1.1	1.6	0.1	0.2
Capillarys 3 Tera IFCC mmol/mol	67.2	5	1.1	1.6	0.8	1.1
Cobas 6000/8000 c501/502 IFCC mmol/mol	67.0	2	1.4	2.1	0.6	0.8
D-10_A1c_Program IFCC mmol/mol	70.6	3	4.7	6.6	4.2	6.3
D-100_A1c_Program IFCC mmol/mol	66.3	8	1.6	2.4	-0.1	-0.2
Hb NEXT IFCC mmol/mol	63.4	4	5.6	8.8	-3.0	-4.5
Minicap flex-piercing IFCC mmol/mol	67.0	4	3.6	5.3	0.6	0.8
TOSOH G11 IFCC mmol/mol	67.2	19	0.8	1.1	0.7	1.1
TOSOH G8 IFCC mmol/mol	66.9	24	1.2	1.7	0.5	0.7

Uw individueel rapport kunt u opvragen via de site  
[https://qseenet.com/CueSee\\_frame.asp?SelectMenu=2](https://qseenet.com/CueSee_frame.asp?SelectMenu=2)

De enquêtecoördinatoren  
 Y. Lenga

Dr. K. Vernelen