

Brussel, 12 juni 2024

Kwaliteit van Laboratoria

GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2024.3D

Target HbA1c = 34.8 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180T IFCC mmol/mol	36.3	3	1.2	3.2	1.6	4.5
ADAMS HA-8180V IFCC mmol/mol	34.1	22	1.1	3.2	-0.7	-1.9
ADAMS HA-8190V IFCC mmol/mol	36.0	2	1.4	3.9	1.2	3.6
Afinion AS100 IFCC mmol/mol	34.0	1	****	****	-0.8	-2.2
Alinity C IFCC mmol/mol	31.7	6	0.4	1.4	-3.0	-8.7
Capillarys 2 flex – piercing IFCC mmol/mol	33.4	4	0.7	2.1	-1.4	-4.0
Capillarys 3 Octa IFCC mmol/mol	33.5	12	1.0	3.0	-1.3	-3.6
Capillarys 3 Tera IFCC mmol/mol	33.4	5	0.9	2.7	-1.4	-3.9
Cobas 6000/8000 c501/502 IFCC mmol/mol	34.5	2	2.1	6.1	-0.3	-0.7
D-10_A1c_Program IFCC mmol/mol	34.5	3	0.5	1.5	-0.3	-0.8
D-100_A1c_Program IFCC mmol/mol	32.9	8	1.4	4.3	-1.9	-5.5
Hb NEXT IFCC mmol/mol	34.3	4	1.9	5.4	-0.5	-1.3
Minicap flex-piercing IFCC mmol/mol	34.5	4	2.4	6.9	-0.3	-0.7
TOSOH G11 IFCC mmol/mol	34.9	19	0.4	1.1	0.2	0.4
TOSOH G8 IFCC mmol/mol	35.2	24	1.0	2.8	0.4	1.3

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=2

De enquêtecoördinatoren
 Y. Lenga

Dr. K. Vernelen