

Brussel, 27 januari 2025

**Kwaliteit van Laboratoria**

**GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2024.6C**

**Target HbA1c = 79.9 mmol/mol (Referentiemethode NGSP/DCCT)**

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180T IFCC mmol/mol	81.0	2	2.82	3.5	1.1	1.4
ADAMS HA-8180V IFCC mmol/mol	80.0	22	2.18	2.7	0.1	0.1
ADAMS HA-8190V IFCC mmol/mol	81.0	1	****	****	1.1	1.4
Afinion 2 IFCC mmol/mol	81.0	1	****	****	1.1	1.4
Afinion AS100 IFCC mmol/mol	79.0	1	****	****	-0.9	-1.1
Alinity C IFCC mmol/mol	77.0	1	****	****	-2.9	-3.6
Capillarys 2 flex-piercing IFCC mmol/mol	77.8	5	2.48	3.2	-2.1	-2.6
Capillarys 3 Octa IFCC mmol/mol	78.3	9	1.66	2.1	-1.6	-1.9
Capillarys 3 Tera IFCC mmol/mol	78.5	6	1.52	1.9	-1.4	-1.7
cobas b 101 IFCC mmol/mol	77.0	1	****	****	-2.9	-3.6
cobas c 501/502 cobas 6000 8000 IFCC	79.0	2	1.41	1.8	-0.9	-1.1
D10 A1c IFCC mmol/mol	82.7	2	3.32	4.0	2.8	3.5
D100 A1c IFCC mmol/mol	81.4	7	1.11	1.4	1.5	1.9
Hb NEXT IFCC mmol/mol	85.6	5	2.63	3.1	5.7	7.1
Minicap flex-piercing IFCC mmol/mol	77.5	4	2.89	3.7	-2.4	-3.0
TOSOH G11 IFCC mmol/mol	81.7	27	0.68	0.8	1.8	2.2
TOSOH G8 IFCC mmol/mol	81.5	16	0.66	0.8	1.6	2.0

Uw individueel rapport kunt u opvragen via de site  
[https://qseenet.com/CueSee\\_frame.asp?SelectMenu1=2](https://qseenet.com/CueSee_frame.asp?SelectMenu1=2)

De enquêtecoördinatoren  
 Y. Lenga

Dr. K. Vernelen/ A. Vantorre