

Brussel, 27 januari 2025

Kwaliteit van Laboratoria

GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2024.6D

Target HbA1c = 34.4 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180T IFCC mmol/mol	35.5	2	2.12	6.0	1.1	3.1
ADAMS HA-8180V IFCC mmol/mol	33.8	22	1.23	3.7	-0.7	-1.9
ADAMS HA-8190V IFCC mmol/mol	35.0	1	***	***	0.6	1.7
Afinion 2 IFCC mmol/mol	34.0	1	***	***	-0.4	-1.2
Afinion AS100 IFCC mmol/mol	38.0	1	***	***	3.6	10.4
Alinity C IFCC mmol/mol	32.2	1	***	***	-2.2	-6.5
Capillarys 2 flex-piercing IFCC mmol/mol	33.4	5	1.14	3.4	-1.0	-3.0
Capillarys 3 Octa IFCC mmol/mol	33.3	9	1.22	3.7	-1.1	-3.2
Capillarys 3 Tera IFCC mmol/mol	33.8	6	0.75	2.2	-0.6	-1.7
cobas b 101 IFCC mmol/mol	33.0	1	***	***	-1.4	-4.1
cobas c 501/502 cobas 6000 8000 IFCC	32.5	2	2.12	6.5	-1.9	-5.6
D10 A1c IFCC mmol/mol	34.7	2	0.42	1.2	0.3	0.8
D100 A1c IFCC mmol/mol	32.0	7	1.44	4.5	-2.4	-7.0
Hb NEXT IFCC mmol/mol	34.7	5	1.42	4.1	0.3	0.7
Minicap flex-piercing IFCC mmol/mol	33.3	4	0.96	2.9	-1.2	-3.4
TOSOH G11 IFCC mmol/mol	34.5	27	0.49	1.4	0.1	0.2
TOSOH G8 IFCC mmol/mol	34.7	16	0.86	2.5	0.3	0.8

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=2

De enquêtecoördinatoren
 Y. Lenga

Dr. K. Vernelen/ A. Vantorre