

Brussel, 19 Februari 2019

Kwaliteit van Laboratoria

GLBAAL RAPPORT EKE HbA1c VERS BLOED - 2019.1D

Target HbA1c = 83.7 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Capillarys 2 flex – piercing IFCC mmol/mol	84.6	10	1.2	1.4	0.8	1.0
Capillarys 3 Tera IFCC mmol/mol	84.5	2	2.1	2.5	0.8	0.9
Cobas c311/111 IFCC mmol/mol	85.2	1	****	****	1.5	1.8
Cobas 6000/8000 c501/502 IFCC mmol/mol	83.2	5	1.4	1.6	-0.6	-0.7
D-10_A1c_Program IFCC mmol/mol	87.5	4	2.4	2.7	3.8	4.5
D-100_A1c_Program IFCC mmol/mol	81.5	7	2.6	3.2	-2.2	-2.6
Menarini HA-8140 IFCC mmol/mol	85.0	1	****	****	1.3	1.5
Menarini HA-8160 TP IFCC mmol/mol	84.2	4	1.5	1.8	0.5	0.6
Menarini HA-8160 VP IFCC mmol/mol	84.3	18	2.1	2.5	0.6	0.8
Menarini HA-8180T IFCC mmol/mol	84.7	4	1.8	2.2	1.0	1.2
Menarini HA-8180V IFCC mmol/mol	85.1	35	2.2	2.6	1.4	1.7
Minicap flex-piercing mmol/mol	86.2	6	0.7	0.9	2.5	2.9
TOSOH G11 IFCC mmol/mol	83.7	1	****	****	0.0	0.0
TOSOH G7 IFCC mmol/mol	83.0	1	****	****	-0.7	-0.9
TOSOH G8 IFCC mmol/mol	85.7	36	1.0	1.2	2.0	2.4
Variant-II_Turbo_A1c_Program IFCC mmol/mol	84.5	2	0.7	0.8	0.8	0.9

Uw individueel rapport kunt u opvragen via de site
http://www.cuesee.com/CueSee_frame.asp?SelectMenu1=2

De enquêtecoördinatoren
 Y. Lenga

Dr. K. Vernelen