

Brussel, 11 Juni 2019

Kwaliteit van Laboratoria

GLBAAL RAPPORT EKE HbA1c VERS BLOED - 2019.3C

Target HbA1c = 63.4 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Capillarys 2 flex – piercing IFCC mmol/mol	64.3	10	1.4	2.2	0.9	1.4
Capillarys 3 Tera IFCC mmol/mol	63.0	2	0.0	0.0	-0.4	-0.6
Cobas c311/111 IFCC mmol/mol	65.6	1	****	****	2.2	3.5
Cobas 6000/8000 c501/502 IFCC mmol/mol	63.6	5	1.1	1.8	0.2	0.4
D-10_A1c_Program IFCC mmol/mol	67.8	4	1.5	2.2	4.4	6.9
D-100_A1c_Program IFCC mmol/mol	63.0	7	0.7	1.2	-0.4	-0.7
Menarini HA-8140 IFCC mmol/mol	64.0	1	****	****	0.6	1.0
Menarini HA-8160 TP IFCC mmol/mol	63.8	4	1.5	2.4	0.4	0.6
Menarini HA-8160 VP IFCC mmol/mol	64.4	18	1.2	1.9	1.0	1.6
Menarini HA-8180T IFCC mmol/mol	64.2	4	1.5	2.3	0.9	1.4
Menarini HA-8180V IFCC mmol/mol	64.9	36	1.4	2.1	1.6	2.5
Minicap flex-piercing mmol/mol	64.6	7	1.6	2.5	1.2	1.9
TOSOH G11 IFCC mmol/mol	64.2	1	****	****	0.8	1.3
TOSOH G7 IFCC mmol/mol	64.0	1	****	****	0.6	1.0
TOSOH G8 IFCC mmol/mol	65.5	34	1.3	1.9	2.1	3.3
Variant-II_Turbo_A1c_Program IFCC mmol/mol	65.0	2	2.8	4.4	1.6	2.5

Uw individueel rapport kunt u opvragen via de site
http://www.cuesee.com/CueSee_frame.asp?SelectMenu1=2

De enquêtecoördinatoren
 Y. Lenga

Dr. K. Vernelen