

Brussel, 14 December 2020

Kwaliteit van Laboratoria

GBAAL RAPPORT EKE HbA1c VERS BLOED - 2020.6C

Target HbA1c = 53.1 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Afinion AS100 IFCC mmol/mol	50.0	1	****	****	-3.1	-5.9
Alinity C IFCC mmol/mol	54.0	1	****	****	0.9	1.7
Capillarys 2 flex – piercing IFCC mmol/mol	53.1	9	1.4	2.6	0.0	0.0
Capillarys 3 Octa IFCC mmol/mol	54.0	1	****	****	0.9	1.7
Capillarys 3 Tera IFCC mmol/mol	53.0	4	0.8	1.5	-0.1	-0.2
Cobas c311/111 IFCC mmol/mol	54.5	1	****	****	1.4	2.6
Cobas 6000/8000 c501/502 IFCC mmol/mol	53.3	5	0.4	0.8	0.2	0.4
D-10_A1c_Program IFCC mmol/mol	52.0	3	2.6	5.1	-1.1	-2.1
D-100_A1c_Program IFCC mmol/mol	52.1	6	1.6	3.2	-1.0	-1.9
Menarini HA-8160 TP IFCC mmol/mol	53.3	3	0.6	1.1	0.2	0.4
Menarini HA-8160 VP IFCC mmol/mol	53.3	18	1.7	3.2	0.2	0.3
Menarini HA-8180T IFCC mmol/mol	53.5	4	1.0	1.9	0.4	0.7
Menarini HA-8180V IFCC mmol/mol	53.9	32	1.6	2.9	0.7	1.4
Minicap flex-piercing mmol/mol	52.2	5	1.1	2.1	-0.9	-1.7
TOSOH G11 IFCC mmol/mol	53.4	3	0.7	1.2	0.3	0.6
TOSOH G8 IFCC mmol/mol	53.4	32	4.2	7.8	0.3	0.5
Variant-II_Turbo_A1c_Program IFCC mmol/mol	54.0	2	0.0	0.0	0.9	1.7

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=2

De enquêtecoördinatoren
 Y. Lenga

Dr. K. Vernelen