

Brussel, 14 December 2020

Kwaliteit van Laboratoria

GLBAAL RAPPORT EKE HbA1c VERS BLOED - 2020.6D

Target HbA1c = 65.0 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Afinion AS100 IFCC mmol/mol	60.0	1	****	****	-5.0	-7.7
Alinity C IFCC mmol/mol	66.0	1	****	****	1.0	1.5
Capillarys 2 flex – piercing IFCC mmol/mol	65.4	9	1.4	2.2	0.4	0.6
Capillarys 3 Octa IFCC mmol/mol	65.0	1	****	****	0.0	0.0
Capillarys 3 Tera IFCC mmol/mol	65.0	4	0.8	1.3	0.0	0.0
Cobas c311/111 IFCC mmol/mol	66.1	3	****	****	1.1	1.7
Cobas 6000/8000 c501/502 IFCC mmol/mol	64.8	5	1.2	1.8	-0.3	-0.4
D-10_A1c_Program IFCC mmol/mol	66.3	3	2.1	3.1	1.3	2.0
D-100_A1c_Program IFCC mmol/mol	64.4	6	1.4	2.2	-0.6	-0.9
Menarini HA-8160 TP IFCC mmol/mol	64.7	3	0.6	0.9	-0.4	-0.6
Menarini HA-8160 VP IFCC mmol/mol	65.0	18	1.9	2.9	-0.1	-0.1
Menarini HA-8180T IFCC mmol/mol	64.6	4	0.7	1.2	-0.4	-0.6
Menarini HA-8180V IFCC mmol/mol	65.5	32	1.6	2.5	0.4	0.7
Minicap flex-piercing mmol/mol	65.4	5	1.1	1.7	0.4	0.6
TOSOH G11 IFCC mmol/mol	66.7	3	0.8	1.1	1.6	2.5
TOSOH G8 IFCC mmol/mol	67.0	32	3.6	5.4	2.0	3.0
Variant-II_Turbo_A1c_Program IFCC mmol/mol	68.0	2	1.4	2.1	3.0	4.6

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=2

De enquêtecoördinatoren
 Y. Lenga

Dr. K. Vernelen