

BIOLOGICAL HEALTH RISKS QUALITY OF LABORATORIES

CLINICAL BIOLOGY COMMISSION

EXTERNAL QUALITY ASSESSMENT IN CLINICAL BIOLOGIE

PRELIMINARY REPORT **Molecular Microbiology HCV-HBV Pilot Study SURVEY 2022/6**

Expected results:

HCV

Sample ID	Matrix	Content	Qualitative result	Quantitative result* (Log IU/mL)
HCV22-1	plasma	HCV	Positive	3.2
HCV22-2	plasma	HCV	Positive	3.83
HCV22-3	plasma	HCV	Positive	3.39
HCV22-4	plasma	HCV	Positive	3.7
HCV22-5	plasma	No HCV	Negative	Not applicable

^{*:} it is the value obtained by the NRC during the validation of the panel

HBV

Sample ID	Matrix	Content	Qualitative result	Quantitative result* (Log IU/mL)
HBV22-1	plasma	HBV	Positive	5.44
HBV22-2	plasma	HBV	Positive	5.26
HBV22-3	plasma	HBV	Positive	4.62
HBV22-4	plasma	HBV	Positive	4.07
HBV22-5	plasma	No HBV	Negative	Not applicable

^{*:} it is the value obtained by the NRC during the validation of the panel

Authorization of the report: By Bernard China, scheme coordinator

Sciensano - Service Quality of laboratories

J. Wytsmanstreet, 14 | 1050 Brussels

Tel.: 02/6425385 E-mail: bernard.china@sciensano.be.

Date of publication: 06/10/2022

© Sciensano Brussels 2022.

This report may not be reproduced, published or distributed without the consent of Sciensano. The laboratories' individual results are confidential. They are not passed on by Sciensano to third parties, nor to members of the Commission, the committees of experts or the working group EQA.

END

