

**BIOLOGICAL HEALTH RISKS
QUALITY OF LABORATORIES**

**PRELIMINARY REPORT
Molecular Microbiology
High-Risk HPV
Pilot Study SURVEY 2023/S4**

Expected results :

Sample ID	Matrix	Content	Expected detected result	Encoded results
HPV23-1	Transport medium	HPV18 HPV16	Positive	93% positive results
HPV23-2	Transport medium	No cellular DNA No HPV	Invalid (no DNA)	100% of negative/ not determined/invalid/inhibition results
HPV23-3	Transport medium	HPV18	Positive	100% positive results
HPV23-4	Transport medium	HPV6	Positive or Negative*	11% positive results 89% negative results
HPV23-5	Transport medium	HPV33	Positive	100% positive results
HPV23-6	Transport medium	HPV53	Positive or Negative*	11% positive results 89% negative results
HPV23-7	Transport medium	HPV16 HPV45	Positive	100% positive results
HPV23-8	Transport medium	HPV51 HPV52 HPV56 HPV58 HPV59	Positive	100% positive results
HPV23-9	Transport medium	HPV39	Positive	96% positive results
HPV23-10	Transport medium	Cellular DNA No HPV	Negative	100% negative results

*Depending on whether or not your method detects low-risk types

Number of datasets: 46

Date of sending: 27/06/2023

Deadline of the encoding of the results: 20/07/2023

Authorization of the report: By Bernard China, scheme coordinator
Sciensano – Service Quality of laboratories
J. Wytmanstreet, 14 | 1050 Brussels
Tel.: 02/6425385 E-mail : bernard.china@sciensano.be.

Date of publication: 31/07/2023

©Sciensano Brussels 2023.

This report may not be reproduced, published or distributed without the consent of Sciensano. The laboratories' individual results are confidential. They are not passed on by Sciensano to third parties, nor to members of the Commission, the committees of experts or the working group EQA.

END